

**FACTORY BATCH CALIBRATION CERTIFICATE  
 FOR LINESCALE-3i ELECTRONIC LOAD CELL**


**Certificate No: 03/24-009**

Capacity: 30 kN / 3000 kgf / 6600 lbf  
 Resolution: 0.01 kN / 1 kgf / 2 lbf  
 Measuring range: -10% to 110% WLL  
 Calibration unit: kgf  
 Acceleration value: 9.807 N/kg  
 Calibration ambient temp: 21° C



The LineScale-3i Batch-9 units with manufacturing date March 2024

- identified with label  
**DOM 03/24 • Batch 009**  
 on the rear plate -



are calibrated to meet the specified load precision

*typical ±0.25% @3kN-15kN (sweet spot)  
 max. ±0.75% @0kN-30kN (full range)*

via the 7-point linear calibration procedure in the range of 0.00 to 30.00 kN.

April 10. 2024

  
 Andreas Riedrich  
 - CEO & Development engineer

  
**LineGrip Corporation Limited**  
 Hauptstr. 41 • 67714 Waldfishbach  
 +49-6333-6030022 • info@linegrip.com

